

GEOSTATS PTY LTD
Mining Industry Consultants
Reference Material Manufacture and Sales

Certified Ore Grade Base Metal Reference Material Product Code

GBM905-11

Certified Control Values

Ore Grade Base Metal Analyses

Element	Grade	Standard Deviation	No of Analyses	Confidence Interval
Nickel (ppm)	38	8	16	+/- 4
Copper (ppm)	31758	1017	47	+/- 302
Zinc (ppm)	84	10	20	+/- 5
Lead (ppm)	42	19	16	+/- 11
Cobalt (ppm)	nr	nr	nr	nr
Silver (ppm)	nr	nr	nr	nr
Sulphur (%)	nr	nr	nr	nr

CRM Details

Control Statistic Details

Control statistics were produced from results accumulated in the :
October-2005 Geostats Pty Ltd Laboratory Round Robin Program.
16 laboratories (at least) tested this material for base metal content.

Source Material

Prior to homogenisation and testing, this material was sourced from
Copper sulphide ore

Colour Designation

Medium light gray

Usage

This product is for use in the mining industry as reference materials for monitoring and testing the accuracy of laboratory assaying.

Preparation and Packaging


All standards are dried in an oven for a minimum of 12 hours at 110C. The dry material is then pulverised to better than 75 micron (nominal mean of 45 micron) using an Air Classifier. The material is then homogenised and stored in a sealed, stable container ready for final packaging.

Materials are statistically sampled from stores, then packaged into either heat sealed, air tight, plastic pulp packets or screw top sealed plastic containers ready for distribution. All packaging has been chosen to ensure minimal contamination from outside sources during shipment, use and storage.

Assay Testwork

All standards are tested thoroughly in the Geostats bi-annual laboratory survey. This involves assaying by a minimum of 50 reputable laboratories selected from across the world using a variety of methods (including AR, 3AD, 4AD and ICP, AAS and XRF). Results are compiled into a comprehensive report detailing statistics for each standard. Assay distributions are checked and processed statistically, producing monitoring statistics for these standards. Materials are tested regularly to ensure stability and homogeneity.

<u>Neutron Activation Analysis Results (ppm)</u>		<u>Major Elements Fusion / XRF (%)</u>	
Antimony	0.4	Fe	nr
Arsenic	3.4	SiO ₂	nr
Barium	<50	Al ₂ O ₃	nr
Bromine	<0.5	TiO ₂	nr
Cadmium	<5	MnO	nr
Cerium	8	CaO	nr
Caesium	0.5	P	nr
Chromium	53	S	nr
Cobalt	43	MgO	nr
Europium	<1	K ₂ O	nr
Gold ppb	214	Na ₂ O	nr
Hafnium	<1	LOI1000	nr
Iridium ppb	<50		
Iron %	5.2		
Lanthanum	4		
Lutetium	<0.2		
Molybdenum	247		
Nickel	30		
Rubidium	18		
Samarium	0.6		
Scandium	0.9		
Selenium	17		
Sodium %	0.14		
Tantalum	<0.5		
Tellurium	<10		
Terbium	<0.5		
Thorium	5.4		
Tin	<100		
Tungsten	1		
Uranium	1		
Ytterbium	<2		
Zinc	<100		
Zirconium	<200		
Calcium%	nr		
Potassium %	nr		
Silver	7		
Mercury	nr		
Neodymium	nr		
Strontium	nr		



10A Marsh Close, O'Connor, Western Australia 6163
Phone : +61 8 9314 2566, Fax : +61 8 9314 3699
e-mail : pjh@geostats.com.au, srr@geostats.com.au
Website <http://www.geostats.com.au>

GBM905-11

Geostats Pty Ltd, Certified Base Metal Reference Material, Product Code :