



Santos Limited
GPO Box 2319
ADELAIDE SA 5000
Australia

Attention: Noel Dowdell

Project **08PEAD0018640**
Client Ref: 849485-101

Customer Sample ID	East Mereenie 43
Sample Type	Gas
Date Sampled	03/06/2008
Time Sampled	1400 h
Pressure	1819 kPag
Temperature	29°C
Cylinder ID	EM43
Test/Reference	Unit

Hydrotesting Cyl	
Hydrotesting*	Pass

GAS ANALYSIS

Test/Reference	Unit
----------------	------

Gas Analysis ASTM D 1945-96 (modified)

Nitrogen	Mol %	6.67
Carbon Dioxide	Mol %	0.09
Methane	Mol %	71.06
Ethane	Mol %	13.53
Propane	Mol %	5.15
I-Butane	Mol %	0.52
N-Butane	Mol %	1.47
I-Pentane	Mol %	0.38
N-Pentane	Mol %	0.46
Hexanes	Mol %	0.39
Heptanes	Mol %	0.19
Octanes and higher hydrocarbons	Mol %	0.09
Total	Mol %	100.00

Gas Parameters ASTM D 1945-96 (modified)

Average Molecular Weight		22.05
Lower Flammability Limit		4.17
Upper Flammability Limit		14.84
Ratio Of Upper To Lower		3.56
Wobbe Index		52.21
Compressibility Factor (Z)		0.9966
Ideal Gas Density (Rel to Air = 1)		0.761
Real Gas Density (Rel to Air = 1)		0.763
Ideal Nett Calorific Value	MJ/m ³	41.35
Ideal Gross Calorific Value	MJ/m ³	45.55
Real Nett Calorific Value	MJ/m ³	41.49
Real Gross Calorific Value	MJ/m ³	45.71
Gross Calorific Val Water-Sat Gas	MJ/m ³	44.76

Test Description

Gas Parameters

The above results are calculated on an air and water free basis assuming only the measured constituents are present. The following parameters are calculated from the above composition at 15°C and 101.325 kPa (abs) using ISO 6976 and the physical constants from the GPSA SI Engineering Data Handbook 11 th Ed.

Authorised By

Neil Thomas

Chemist

Michelle Fordham

Chemist

Accreditation No 2013

Laboratory Manager

Diane Cass

Operations Manager

**Final Report**

- Indicates Not Requested

* Indicates NATA accreditation does not cover the performance of this service

Samples will be discarded after 30 days unless otherwise notified.

Amdel Limited shall not be liable for loss, cost, damages or expenses incurred by the client, or any other person or company, resulting from the use of any information or interpretation given in this report. In no case shall Amdel Limited be liable for consequential damages including, but not limited to, lost profits, damages for failure to meet deadlines and lost production arising from this report. This document shall not be reproduced except in full and relates only to the items tested. Unless indicated otherwise, the tests were performed on the samples as received. *The samples were not collected by Amdel staff.*



NATIONAL HYDRO TESTING SERVICES
 1 ELLEMSEA CCT. LONSDALE SA 5160
 PH 8326 7755 MOBILE 0418 804 144

REGISTERED

MARK

REF

#AAO 284 (EM 43)

Date of Test:	8/08
Previous Test Date:	
Reference Temperature:	14
Service:	I

TESTED IN COMPLIANCE WITH AS2030.1

CERTIFICATE OF TEST 29587

S.A.A. APPROVED

TEST STATION No.: 308A

Client: AMDEL

Serial No.: AAO 284

Specification No.: DOT 3E-1800

Water Capacity: .5L

Test Pressure: 2925 PSI

Working Pressure: 1800 PSI

Condition: External

Condition: Internal

HYDROSTATIC/VISUAL TEST RESULT:

Passed

WEIGHT TEST: Passed

Good

Good

Recommendation or Remarks:

Average

Average

Poor

Poor

Approved Signatory: 