



Australian Laboratory Services Pty. Ltd.
32 Shand Street
Stafford
Brisbane QLD 4053
Phone: +61 (7) 3243 7222 Fax: +61 (7) 3243 7218
www.alsglobal.com

Page: 1
Total # Pages: 3 (A - E)
Plus Appendix Pages
Finalized Date: 18-MAY-2015
This copy reported on 20-MAY-2015
Account: FINCON

QC CERTIFICATE PH15068596

P.O. No.: 308773

This report is for 2 Rock samples submitted to our lab in Perth, WA, Australia on 8-MAY-2015.

The following have access to data associated with this certificate:

KIM MURRAY

SAMPLE PREPARATION

ALS CODE	DESCRIPTION
LOG-22	Sample login - Rcd w/o BarCode
CRU-21	Crush entire sample >70% -6 mm
PUL-21	Pulverize entire sample
WEI-21	Received Sample Weight
LEV-01	Waste Disposal Levy
BAG-01	Bulk Master for Storage
PUL-QC	Pulverizing QC Test
SND-01	Send out to other laboratory

ANALYTICAL PROCEDURES

ALS CODE	DESCRIPTION	INSTRUMENT
ME-GRA05	H2O/LOI by TGA furnace	TGA
ME-MS61	48 element four acid ICP-MS	
ME-XRF21n	Iron Ore by XRF Fusion	XRF

To: FINNIS CONTRACTING PTY LTD
ATTN: KIM MURRAY
590 LITCHFIELD PARK ROAD
BATCHELOR NT 0845

This is the Final Report and supersedes any preliminary report with this certificate number. Results apply to samples as submitted. All pages of this report have been checked and approved for release.

***** See Appendix Page for comments regarding this certificate *****

Signature:

Cameron Brosnan, Laboratory Manager, Perth



Australian Laboratory Services Pty. Ltd.
 32 Shand Street
 Stafford
 Brisbane QLD 4053
 Phone: +61 (7) 3243 7222 Fax: +61 (7) 3243 7218
 www.alsglobal.com

Page: 2 - A
 Total # Pages: 3 (A - E)
 Plus Appendix Pages
 Finalized Date: 18-MAY-2015
 Account: FINCON

QC CERTIFICATE OF ANALYSIS PH15068596

Method Analyte Units LOR	ME-XRF21n Al2O3 %	ME-XRF21n As %	ME-XRF21n Ba %	ME-XRF21n CaO %	ME-XRF21n Cl %	ME-XRF21n Co %	ME-XRF21n Cr2O3 %	ME-XRF21n Cu %	ME-XRF21n Fe %	ME-XRF21n K2O %	ME-XRF21n MgO %	ME-XRF21n Mn %	ME-XRF21n Na2O %	ME-XRF21n Ni %	ME-XRF21n P %
Sample Description	0.01	0.001	0.001	0.01	0.001	0.001	0.001	0.001	0.01	0.001	0.01	0.001	0.005	0.001	0.001
STANDARDS															
GBM908-10															
Target Range - Lower Bound															
Upper Bound															
MRGeo08															
Target Range - Lower Bound															
Upper Bound															
NCSDC18014	1.32	0.003	0.026	0.16	0.011	0.002	0.002	0.002	65.92	0.200	0.03	0.033	0.015	0.002	0.072
Target Range - Lower Bound	1.21	<0.001	0.022	0.12	0.011	<0.001	<0.001	0.002	65.13	0.186	0.03	0.030	0.012	<0.001	0.066
Upper Bound	1.39	0.005	0.026	0.18	0.017	0.004	0.006	0.006	66.61	0.208	0.07	0.035	0.034	0.004	0.076
OREAS 24b															
Target Range - Lower Bound															
Upper Bound															
OREAS 403	2.63	0.001	<0.001	0.11	0.005	0.001	0.010	0.002	52.39	0.012	0.08	0.004	0.011	<0.001	0.122
Target Range - Lower Bound	2.49	<0.001	<0.001	0.08	<0.001	<0.001	0.008	<0.001	51.68	0.009	0.05	0.003	<0.005	<0.001	0.116
Upper Bound	2.77	0.003	0.002	0.13	0.006	0.002	0.013	0.004	52.93	0.013	0.10	0.007	0.026	0.004	0.130
ST-392															
Target Range - Lower Bound															
Upper Bound															
BLANKS															
BLANK															
Target Range - Lower Bound															
Upper Bound															
BLANK	0.02	<0.001	<0.001	0.01	0.001	0.002	<0.001	0.002	<0.01	0.001	0.01	<0.001	0.006	0.002	<0.001
Target Range - Lower Bound	<0.01	<0.001	<0.001	<0.01	<0.001	<0.001	<0.001	<0.001	<0.01	<0.001	<0.01	<0.001	<0.005	<0.001	<0.001
Upper Bound	0.02	0.002	0.002	0.02	0.002	0.002	0.002	0.002	0.02	0.002	0.02	0.002	0.010	0.002	0.002



Australian Laboratory Services Pty. Ltd.
32 Shand Street
Stafford
Brisbane QLD 4053
Phone: +61 (7) 3243 7222 Fax: +61 (7) 3243 7218
www.alsglobal.com

Page: 2 - B
Total # Pages: 3 (A - E)
Plus Appendix Pages
Finalized Date: 18-MAY-2015
Account: FINCON

QC CERTIFICATE OF ANALYSIS PH15068596

Method Analyte Units LOR	ME-XRF21n Pb %	ME-XRF21n S %	ME-XRF21n SiO2 %	ME-XRF21n Sn %	ME-XRF21n Sr %	ME-XRF21n TiO2 %	ME-XRF21n V %	ME-XRF21n Zn %	ME-XRF21n Zr %	ME-XRF21n Total %	ME-GRA05 LOI %	ME-MS61 Ag ppm	ME-MS61 Al %	ME-MS61 As ppm	ME-MS61 Ba ppm
Sample Description	0.001	0.001	0.01	0.001	0.001	0.01	0.001	0.001	0.001	0.01	0.01	0.01	0.01	0.2	10
STANDARDS															
GBM908 - 10												2.86	6.88	55.6	1110
Target Range - Lower Bound												2.69	6.40	51.1	930
Upper Bound												3.31	7.84	62.9	1280
MRGeo08												4.03	7.12	31.6	1060
Target Range - Lower Bound												4.00	6.64	29.7	920
Upper Bound												4.92	8.14	36.7	1270
NCSDC18014	0.006	0.021	3.15	0.001	0.025	0.06	0.003	0.003	0.004	100.05					
Target Range - Lower Bound	0.005	0.018	2.99	<0.001	0.020	0.03	<0.001	<0.001	<0.001	97.99					
Upper Bound	0.010	0.022	3.31	0.002	0.026	0.07	0.004	0.005	0.005	102.00					
OREAS 24b												0.06	7.23	8.3	700
Target Range - Lower Bound												0.08	7.02	7.3	600
Upper Bound												0.13	8.60	9.4	830
OREAS 403	0.003	0.028	13.55	<0.001	0.001	0.29	0.003	0.001	0.009	100.00					
Target Range - Lower Bound	<0.001	0.024	13.20	<0.001	<0.001	0.26	<0.001	<0.001	0.006	97.99					
Upper Bound	0.004	0.028	14.15	0.002	0.002	0.31	0.006	0.004	0.011	102.00					
ST-392											5.52				
Target Range - Lower Bound											5.35				
Upper Bound											5.71				
BLANKS															
BLANK												0.01	<0.01	<0.2	<10
Target Range - Lower Bound												<0.01	<0.01	<0.2	<10
Upper Bound												0.02	0.02	0.4	20
BLANK	0.006	<0.001	99.9	<0.001	<0.001	<0.01	<0.001	0.001	<0.001	99.96					
Target Range - Lower Bound	<0.001	<0.001	98.0	<0.001	<0.001	<0.01	<0.001	<0.001	<0.001	97.99					
Upper Bound	0.002	0.002	100.0	0.002	0.002	0.02	0.002	0.002	0.002	102.00					



Australian Laboratory Services Pty. Ltd.
 32 Shand Street
 Stafford
 Brisbane QLD 4053
 Phone: +61 (7) 3243 7222 Fax: +61 (7) 3243 7218
 www.alsglobal.com

Page: 2 - C
 Total # Pages: 3 (A - E)
 Plus Appendix Pages
 Finalized Date: 18-MAY-2015
 Account: FINCON

QC CERTIFICATE OF ANALYSIS PH15068596

Method Analyte Units LOR	ME-MS61 Be ppm 0.05	ME-MS61 Bi ppm 0.01	ME-MS61 Ca % 0.01	ME-MS61 Cd ppm 0.02	ME-MS61 Ce ppm 0.01	ME-MS61 Co ppm 0.1	ME-MS61 Cr ppm 1	ME-MS61 Cs ppm 0.05	ME-MS61 Cu ppm 0.2	ME-MS61 Fe % 0.01	ME-MS61 Ga ppm 0.05	ME-MS61 Ge ppm 0.05	ME-MS61 Hf ppm 0.1	ME-MS61 In ppm 0.005	ME-MS61 K % 0.01
Sample Description															
STANDARDS															
GBM908-10	1.41	1.15	3.76	1.83	103.5	25.0	142	3.82	3650	5.65	22.0	0.28	3.9	0.075	2.15
Target Range - Lower Bound	1.27	1.12	3.33	1.53	99.0	23.3	125	3.44	3380	4.98	18.65	0.18	3.3	0.064	1.87
Upper Bound	1.66	1.39	4.10	1.91	121.0	28.7	155	4.32	3880	6.10	22.9	0.40	4.3	0.092	2.31
MRGeo08	2.95	0.67	2.55	2.21	67.5	18.0	91	11.55	608	3.96	19.40	0.18	3.2	0.169	3.09
Target Range - Lower Bound	2.91	0.62	2.35	2.05	66.7	17.7	81	11.00	587	3.55	17.50	<0.05	2.8	0.161	2.79
Upper Bound	3.67	0.78	2.90	2.55	81.5	21.9	102	13.60	675	4.37	21.5	0.27	3.6	0.207	3.43
NCSDC18014															
Target Range - Lower Bound															
Upper Bound															
OREAS 24b	2.32	1.02	1.03	0.06	83.6	15.3	114	10.30	38.7	4.28	21.5	0.20	4.2	0.073	2.70
Target Range - Lower Bound	2.25	0.60	0.94	<0.02	76.8	15.1	105	9.58	35.1	3.94	18.05	<0.05	3.4	0.063	2.52
Upper Bound	2.86	0.76	1.18	0.09	93.8	18.7	131	11.80	40.9	4.84	22.2	0.26	4.4	0.091	3.10
OREAS 403															
Target Range - Lower Bound															
Upper Bound															
ST-392															
Target Range - Lower Bound															
Upper Bound															
BLANKS															
BLANK	<0.05	<0.01	<0.01	<0.02	<0.01	<0.1	<1	<0.05	0.3	<0.01	<0.05	<0.05	<0.1	<0.005	<0.01
Target Range - Lower Bound	<0.05	<0.01	<0.01	<0.02	<0.01	<0.1	<1	<0.05	<0.2	<0.01	<0.05	<0.05	<0.1	<0.005	<0.01
Upper Bound	0.10	0.02	0.02	0.04	0.02	0.2	2	0.10	0.4	0.02	0.10	0.10	0.2	0.010	0.02
BLANK															
Target Range - Lower Bound															
Upper Bound															



Australian Laboratory Services Pty. Ltd.
 32 Shand Street
 Stafford
 Brisbane QLD 4053
 Phone: +61 (7) 3243 7222 Fax: +61 (7) 3243 7218
 www.alsglobal.com

Page: 2 - D
 Total # Pages: 3 (A - E)
 Plus Appendix Pages
 Finalized Date: 18-MAY-2015
 Account: FINCON

QC CERTIFICATE OF ANALYSIS PH15068596

Method Analyte Units LOR	ME-MS61 La ppm 0.5	ME-MS61 Li ppm 0.2	ME-MS61 Mg % 0.01	ME-MS61 Mn ppm 5	ME-MS61 Mo ppm 0.05	ME-MS61 Na % 0.01	ME-MS61 Nb ppm 0.1	ME-MS61 Ni ppm 0.2	ME-MS61 P ppm 10	ME-MS61 Pb ppm 0.5	ME-MS61 Rb ppm 0.1	ME-MS61 Re ppm 0.002	ME-MS61 S % 0.01	ME-MS61 Sb ppm 0.05	ME-MS61 Sc ppm 0.1
Sample Description															
STANDARDS															
GBM908 - 10	52.5	10.8	1.78	799	68.2	2.12	11.1	2230	990	2080	163.5	<0.002	0.37	1.78	17.8
Target Range - Lower Bound	49.0	9.8	1.61	715	57.9	1.93	9.5	2030	880	1860	153.0	<0.002	0.33	1.40	17.0
Upper Bound	61.0	12.4	1.99	885	70.9	2.38	11.9	2480	1100	2270	187.0	0.006	0.43	2.01	21.0
MRGeo08	32.6	32.5	1.28	537	15.00	1.89	20.9	677	1040	1060	185.5	0.007	0.29	4.55	11.2
Target Range - Lower Bound	31.1	30.4	1.17	497	13.65	1.76	19.0	622	930	971	173.5	0.006	0.27	3.89	11.1
Upper Bound	39.1	37.6	1.45	619	16.75	2.18	23.4	760	1160	1185	212	0.016	0.36	5.39	13.7
NCSDC18014															
Target Range - Lower Bound															
Upper Bound															
OREAS 24b	41.8	55.3	1.49	422	3.82	0.77	15.2	69.8	670	22.9	163.5	<0.002	0.19	1.01	14.8
Target Range - Lower Bound	37.7	48.4	1.39	391	3.46	0.70	13.0	53.8	610	20.3	147.5	<0.002	0.17	0.80	13.7
Upper Bound	47.1	59.6	1.72	489	4.34	0.88	16.2	66.2	770	25.9	180.5	0.006	0.23	1.20	16.9
OREAS 403															
Target Range - Lower Bound															
Upper Bound															
ST-392															
Target Range - Lower Bound															
Upper Bound															
BLANKS															
BLANK	<0.5	<0.2	<0.01	<5	<0.05	<0.01	<0.1	<0.2	<10	<0.5	<0.1	<0.002	0.01	<0.05	<0.1
Target Range - Lower Bound	<0.5	<0.2	<0.01	<5	<0.05	<0.01	<0.1	<0.2	<10	<0.5	<0.1	<0.002	<0.01	<0.05	<0.1
Upper Bound	1.0	0.4	0.02	10	0.10	0.02	0.2	0.4	20	1.0	0.2	0.004	0.02	0.10	0.2
BLANK															
Target Range - Lower Bound															
Upper Bound															

QC CERTIFICATE OF ANALYSIS PH15068596

Method Analyte Units LOR	ME-MS61 Se ppm 1	ME-MS61 Sn ppm 0.2	ME-MS61 Sr ppm 0.2	ME-MS61 Ta ppm 0.05	ME-MS61 Te ppm 0.05	ME-MS61 Th ppm 0.2	ME-MS61 Ti % 0.005	ME-MS61 Tl ppm 0.02	ME-MS61 U ppm 0.1	ME-MS61 V ppm 1	ME-MS61 W ppm 0.1	ME-MS61 Y ppm 0.1	ME-MS61 Zn ppm 2	ME-MS61 Zr ppm 0.5
Sample Description														
STANDARDS														
GBM908-10	2	3.4	291	0.75	0.05	16.8	0.669	1.19	2.2	143	3.4	37.2	1090	143.0
Target Range - Lower Bound	<1	2.7	258	0.68	<0.05	16.4	0.591	1.00	2.0	123	2.9	35.2	939	117.5
Upper Bound	4	3.9	316	0.97	0.16	20.4	0.733	1.40	2.6	153	4.1	43.2	1155	160.5
MRGeo08	2	4.0	299	1.50	<0.05	18.1	0.494	1.01	5.2	109	4.8	24.7	790	105.5
Target Range - Lower Bound	<1	3.5	277	1.39	<0.05	17.7	0.454	0.89	4.9	97	4.1	23.8	722	92.2
Upper Bound	4	4.7	339	1.81	0.14	22.1	0.566	1.25	6.3	121	5.8	29.3	886	126.0
NCSDC18014														
Target Range - Lower Bound														
Upper Bound														
OREAS 24b	1	4.8	122.0	1.15	<0.05	16.3	0.436	0.85	3.0	107	3.9	18.9	106	139.0
Target Range - Lower Bound	<1	3.6	111.5	1.06	<0.05	14.6	0.387	0.71	2.7	96	3.0	18.3	93	113.5
Upper Bound	3	4.9	136.5	1.40	0.10	18.2	0.484	1.01	3.5	120	4.3	22.5	118	154.5
OREAS 403														
Target Range - Lower Bound														
Upper Bound														
ST-392														
Target Range - Lower Bound														
Upper Bound														
BLANKS														
BLANK	<1	<0.2	<0.2	<0.05	<0.05	<0.2	<0.005	<0.02	<0.1	<1	<0.1	<0.1	2	<0.5
Target Range - Lower Bound	<1	<0.2	<0.2	<0.05	<0.05	<0.2	<0.005	<0.02	<0.1	<1	<0.1	<0.1	<2	<0.5
Upper Bound	5	0.4	0.4	0.10	0.10	0.4	0.010	0.04	0.2	2	0.2	0.2	4	1.0
BLANK														
Target Range - Lower Bound														
Upper Bound														



Australian Laboratory Services Pty. Ltd.
32 Shand Street
Stafford
Brisbane QLD 4053
Phone: +61 (7) 3243 7222 Fax: +61 (7) 3243 7218
www.alsglobal.com

Page: 3 - A
Total # Pages: 3 (A - E)
Plus Appendix Pages
Finalized Date: 18-MAY-2015
Account: FINCON

QC CERTIFICATE OF ANALYSIS PH15068596

Sample Description	Method Analyte Units LOR	ME-XRF21n Al2O3 %	ME-XRF21n As %	ME-XRF21n Ba %	ME-XRF21n CaO %	ME-XRF21n Cl %	ME-XRF21n Co %	ME-XRF21n Cr2O3 %	ME-XRF21n Cu %	ME-XRF21n Fe %	ME-XRF21n K2O %	ME-XRF21n MgO %	ME-XRF21n Mn %	ME-XRF21n Na2O %	ME-XRF21n Ni %	ME-XRF21n P %
		0.01	0.001	0.001	0.01	0.001	0.001	0.001	0.001	0.01	0.001	0.01	0.001	0.005	0.001	0.001
KL2 DUP		DUPLICATES														
Target Range - Lower Bound																
Upper Bound																
KL4 DUP		1.17	<0.001	0.004	0.02	0.026	0.002	0.001	0.002	66.00	0.276	0.05	0.026	0.019	0.003	0.020
Target Range - Lower Bound		1.19	<0.001	0.005	0.02	0.029	0.002	0.002	0.002	65.93	0.277	0.06	0.027	0.022	0.003	0.020
Upper Bound		1.16	<0.001	0.003	<0.01	0.025	<0.001	<0.001	<0.001	65.63	0.269	0.04	0.025	0.014	0.002	0.019
		1.20	0.002	0.006	0.03	0.030	0.003	0.002	0.003	66.30	0.284	0.07	0.028	0.027	0.004	0.022



Australian Laboratory Services Pty. Ltd.
 32 Shand Street
 Stafford
 Brisbane QLD 4053
 Phone: +61 (7) 3243 7222 Fax: +61 (7) 3243 7218
 www.alsglobal.com

Page: 3 - B
 Total # Pages: 3 (A - E)
 Plus Appendix Pages
 Finalized Date: 18-MAY-2015
 Account: FINCON

QC CERTIFICATE OF ANALYSIS PH15068596

Sample Description	Method Analyte Units LOR	ME-XRF21n Pb %	ME-XRF21n S %	ME-XRF21n SiO2 %	ME-XRF21n Sn %	ME-XRF21n Sr %	ME-XRF21n TiO2 %	ME-XRF21n V %	ME-XRF21n Zn %	ME-XRF21n Zr %	ME-XRF21n Total %	ME-GRA05 LOI %	ME-MS61 Ag ppm	ME-MS61 Al %	ME-MS61 As ppm	ME-MS61 Ba ppm
		0.001	0.001	0.01	0.001	0.001	0.01	0.001	0.001	0.001	0.01	0.01	0.01	0.01	0.2	10
KL2 DUP		DUPLICATES														
Target Range - Lower Bound												0.35				
Upper Bound												0.35				
KL4 DUP		0.005	0.006	3.48	0.001	0.001	0.08	0.008	0.003	0.003	100.00		0.10	0.55	2.2	20
Target Range - Lower Bound		0.005	0.008	3.52	<0.001	<0.001	0.08	0.008	0.003	0.003	100.00		0.11	0.55	2.4	20
Upper Bound		0.004	0.006	3.46	<0.001	<0.001	0.07	0.007	0.002	0.002	98.99		0.09	0.51	2.0	<10
		0.006	0.008	3.55	0.002	0.002	0.09	0.009	0.004	0.004	101.00		0.12	0.59	2.6	30



Australian Laboratory Services Pty. Ltd.
32 Shand Street
Stafford
Brisbane QLD 4053
Phone: +61 (7) 3243 7222 Fax: +61 (7) 3243 7218
www.alsglobal.com

Page: 3 - C
Total # Pages: 3 (A - E)
Plus Appendix Pages
Finalized Date: 18-MAY-2015
Account: FINCON

QC CERTIFICATE OF ANALYSIS PH15068596

Sample Description	Method Analyte Units LOR	ME-MS61 Be ppm 0.05	ME-MS61 Bi ppm 0.01	ME-MS61 Ca % 0.01	ME-MS61 Cd ppm 0.02	ME-MS61 Ce ppm 0.01	ME-MS61 Co ppm 0.1	ME-MS61 Cr ppm 1	ME-MS61 Cs ppm 0.05	ME-MS61 Cu ppm 0.2	ME-MS61 Fe % 0.01	ME-MS61 Ga ppm 0.05	ME-MS61 Ge ppm 0.05	ME-MS61 Hf ppm 0.1	ME-MS61 In ppm 0.005	ME-MS61 K % 0.01
KL2 DUP Target Range - Lower Bound Upper Bound		DUPLICATES														
KL4 DUP Target Range - Lower Bound Upper Bound		0.67 0.77 0.63 0.81	0.10 0.10 0.09 0.12	0.01 0.01 <0.01 0.02	<0.02 <0.02 <0.02 0.04	12.15 12.05 11.50 12.70	7.3 7.3 6.8 7.8	12 19 14 17	0.18 0.18 0.12 0.24	5.9 5.8 5.4 6.3	41.1 40.9 38.9 43.1	4.56 4.55 4.28 4.83	0.41 0.49 0.38 0.52	0.4 0.4 0.3 0.5	0.034 0.035 0.028 0.041	0.22 0.23 0.20 0.25



Australian Laboratory Services Pty. Ltd.
32 Shand Street
Stafford
Brisbane QLD 4053
Phone: +61 (7) 3243 7222 Fax: +61 (7) 3243 7218
www.alsglobal.com

Page: 3 - D
Total # Pages: 3 (A - E)
Plus Appendix Pages
Finalized Date: 18-MAY-2015
Account: FINCON

QC CERTIFICATE OF ANALYSIS PH15068596

Sample Description	Method Analyte Units LOR	ME-MS61 La ppm 0.5	ME-MS61 Li ppm 0.2	ME-MS61 Mg % 0.01	ME-MS61 Mn ppm 5	ME-MS61 Mo ppm 0.05	ME-MS61 Na % 0.01	ME-MS61 Nb ppm 0.1	ME-MS61 Ni ppm 0.2	ME-MS61 P ppm 10	ME-MS61 Pb ppm 0.5	ME-MS61 Rb ppm 0.1	ME-MS61 Re ppm 0.002	ME-MS61 S % 0.01	ME-MS61 Sb ppm 0.05	ME-MS61 Sc ppm 0.1
KL2 DUP Target Range - Lower Bound Upper Bound		DUPLICATES														
KL4 DUP Target Range - Lower Bound Upper Bound		6.3 6.2 5.4 7.1	0.9 0.9 0.7 1.1	0.02 0.02 <0.01 0.03	148 173 147 174	0.75 0.71 0.64 0.82	0.02 0.02 <0.01 0.03	0.7 0.8 0.6 0.9	21.0 21.1 19.8 22.3	140 140 120 160	3.7 3.5 2.9 4.3	4.3 4.3 4.0 4.6	<0.002 <0.002 <0.002 0.004	0.01 0.01 <0.01 0.02	0.36 0.35 0.28 0.43	4.0 4.1 3.7 4.4



Australian Laboratory Services Pty. Ltd.
 32 Shand Street
 Stafford
 Brisbane QLD 4053
 Phone: +61 (7) 3243 7222 Fax: +61 (7) 3243 7218
 www.alsglobal.com

Page: 3 - E
 Total # Pages: 3 (A - E)
 Plus Appendix Pages
 Finalized Date: 18-MAY-2015
 Account: FINCON

QC CERTIFICATE OF ANALYSIS PH15068596

Sample Description	Method Analyte Units LOR	ME-MS61 Se ppm 1	ME-MS61 Sn ppm 0.2	ME-MS61 Sr ppm 0.2	ME-MS61 Ta ppm 0.05	ME-MS61 Te ppm 0.05	ME-MS61 Th ppm 0.2	ME-MS61 Ti % 0.005	ME-MS61 TI ppm 0.02	ME-MS61 U ppm 0.1	ME-MS61 V ppm 1	ME-MS61 W ppm 0.1	ME-MS61 Y ppm 0.1	ME-MS61 Zn ppm 2	ME-MS61 Zr ppm 0.5
KL2 DUP Target Range - Lower Bound Upper Bound		DUPLICATES													
KL4 DUP Target Range - Lower Bound Upper Bound		<1 <1 <1 2	11.0 10.9 10.2 11.7	3.0 2.9 2.6 3.3	<0.05 <0.05 <0.05 0.10	<0.05 <0.05 <0.05 0.10	0.9 0.8 0.6 1.1	0.033 0.034 0.027 0.040	<0.02 <0.02 <0.02 0.04	1.4 1.4 1.2 1.6	49 49 46 52	2.9 3.1 2.7 3.3	4.8 4.6 4.4 5.0	16 16 13 19	17.7 18.9 16.9 19.7



Australian Laboratory Services Pty. Ltd.
32 Shand Street
Stafford
Brisbane QLD 4053
Phone: +61 (7) 3243 7222 Fax: +61 (7) 3243 7218
www.alsglobal.com

Page: Appendix 1
Total # Appendix Pages: 1
Finalized Date: 18-MAY-2015
Account: FINCON

QC CERTIFICATE OF ANALYSIS PH15068596

	CERTIFICATE COMMENTS
	ANALYTICAL COMMENTS
Applies to Method:	REE's may not be totally soluble in this method. ME-MS61
	LABORATORY ADDRESSES
Applies to Method:	Processed at ALS Perth located at 32 Oxleigh Drive, Malaga, Perth, WA, Australia. BAG-01 CRU-21 LEV-01 LOG-22 ME-MS61 PUL-21 PUL-QC SND-01 WEI-21
Applies to Method:	Processed at ALS Iron Ore Technical Centre located at 26 Rigali Way, Wangara, Perth, WA, Australia. ME-GRA05 ME-XRF21n