

**RADIATION
SERVICES**WORLD RECOGNISED
ACCREDITATIONAccreditation No. 16987
Accredited for compliance
with ISO/IEC 17025

Certificate of Calibration

Certificate No.: CAL-16-12321**A. Issued to**

Company name: SGS Radiation Services
Address: 10 / 585 Blackburn Road
Notting Hill VIC 3168

B. Instrument and use

Manufacturer: Georadis
Model No.: RS 220
Serial No.: 1710
Intended use: Gamma ray survey and radionuclide identification

C. Details of calibration

Previous calibration: 6th October 2015

Date calibrated: 17th October 2016

Type of test: Annual calibration.

Test conditions: Gamma radiation emitting point sources were used to determine the instrument's response to different radionuclides (test method: ARS-SOP-CL6). The response of the instrument to a beam of collimated gamma rays from one or more ¹³⁷Cs sources was also evaluated (test method: ARS-SOP-CL2).

Test reference point: NaI detector: 40 mm behind base.

Orientation: Horizontal in radiation beam.

Uncertainties: The uncertainties stated on this certificate are expressed with a coverage factor of 2 corresponding to a 95% confidence level.

Traceability: The outputs of all ¹³⁷Cs sources used to calibrate this instrument are traceable to the Australian Primary Standard of Exposure maintained by ARPANSA.

D. Calibration results

The instrument over responded by greater than 25% at ambient dose equivalent rates greater than 43.8 $\mu\text{Sv}\cdot\text{h}^{-1}$.

Refer to the results on page 2 of this certificate for the response of the instrument at specified test points.

- Notes: 1. Before using this instrument the user should be familiar with its characteristics (energy dependence, directional dependence, etc.).
2. Any alteration to the electronic settings or damage to the instrument will render the calibration void.

Calibrated by: Mr. Roland Sargent
Position: Senior Technical Officer
Signature:

Reviewed by: Ms. Gillian Nicholson
Position: Senior Technical Officer
Signature:

Date: 17th October 2016

Date: 18th October 2016

Certificate No.: CAL-16-12321

E. Dose response (^{137}Cs – 662 keV)

Radiation source	Instrument range	Measured ambient dose equivalent rate ^a ($\mu\text{Sv}\cdot\text{h}^{-1}$)	Actual ambient dose equivalent rate ($\mu\text{Sv}\cdot\text{h}^{-1}$)	Correction factor ^b	Estimated uncertainty
Background	Auto	0.03 ^c	N/A ^d	N/A	N/A
^{137}Cs (3.7 GBq)	Auto	0.28	0.29	1.04	$\pm 10\%$
^{137}Cs (3.7 GBq)	Auto	1.07	1.12	1.05	$\pm 10\%$
^{137}Cs (3.7 GBq)	Auto	2.05	2.12	1.03	$\pm 9\%$
^{137}Cs (3.7 GBq)	Auto	10.9	11.4	1.04	$\pm 8\%$
^{137}Cs (3.7 GBq)	Auto	47.7	43.8	0.92	$\pm 6\%$

- a. Background correction has been applied.
b. Correction factor is the ratio of the background corrected actual ambient dose equivalent rate to the background corrected measured ambient dose equivalent rate reading.
c. No background correction applies.
d. The actual background has been measured in units of air kerma ($\mu\text{Gy}\cdot\text{h}^{-1}$) using Health Physics Instruments Model 1010 (SN:419).

F. Source identification

Source and actual energies (keV) of characteristic peaks	Measured energies (keV) of characteristic peaks ^a	Characteristic peaks identified	Source identified ^b
^{137}Cs (662)	662	Yes	Yes
^{22}Na (511, 1275)	512, 1273	Yes	No ^c
^{133}Ba (81, 303, 356)	80.5, 302, 356	Yes	Yes
^{60}Co (1173, 1333)	1162, 1320	Yes	Yes
^{152}Eu (344, 779, 1112, 1408)	341, 770, 1091, 1397	Yes	Yes
^{226}Ra (352, 609, 1120)	349, 604, 1111	Yes	Yes
^{241}Am (60)	58.2	Yes	Yes
^{232}Th (239, 338, 583)	240, 344, 589	Yes	Yes

- a. The characteristic peaks are listed as noted from the peak analysis menu of the RS-220.
b. Custom library used for source identification test.
c. Radionuclide not in custom library.

Certificate No.: CAL-16-12321

Appendix A: Reference source specifications

Source No.	Radiation source	Activity	Source type	Serial No.	Reference date
1	¹³⁷ Cs	370 MBq	Calibration Source	GN5479	6 th March 2007
2	¹³⁷ Cs	3.7 GBq	Calibration Source	GH0466	20 th May 2009
3	¹³⁷ Cs	41 GBq	Calibration Source	GN5479	20 th May 2009
4	¹³⁷ Cs	145 GBq	Calibration Source	SG781	16 th November 2009
5	²³⁸ Pu	892 Bq	Wide Area	HA 943	19 th January 2000
6	²³⁸ Pu ^a	3.61 kBq	Wide Area	LT 178	28 th May 2003
7	²⁴¹ Am ^b	2.82 kBq	Wide Area	UD 805	28 th February 2012
8	¹⁴ C	2.50 kBq	Wide Area	HM 470	28 th September 2000
9	⁹⁹ Tc	1.10 kBq	Wide Area	GF 855	15 th October 1998
10	³⁶ Cl	1.22 kBq	Wide Area	GF 854	6 th October 1998
11	⁹⁰ Sr/ ⁹⁰ Y	1.19 kBq	Wide Area	HR 157	10 th January 2001
12	¹²⁹ I	943 Bq	Wide Area	HM 469	16 th October 2000
13	²¹⁰ Pb	223 kBq	Point Source	FZ165	1 st October 1998
14	²⁴¹ Am	36.6 kBq	Point Source	FZ159	1 st October 1998
15	¹³⁷ Cs	40.0 kBq	Point Source	FZ161	1 st October 1998
16	⁶⁰ Co	44.1 kBq	Point Source	FZ162	1 st October 1998
17	²²⁶ Ra	55.7 kBq	Point Source	FZ166	1 st October 1998
18	²² Na	41.0 kBq	Point Source	FZ164	1 st October 1998
19	¹⁵² Eu	44.4 kBq	Point Source	FZ163	1 st October 1998
20	¹³³ Ba	38.9 kBq	Point Source	FZ160	1 st October 1998
21	¹³⁷ Cs	10 kBq	Point Source	SGS10KBq	13 th August 2014
22	¹³⁷ Cs	62 kBq	Point Source	SGS62KBq	13 th August 2014
23	¹³⁷ Cs	335 kBq	Point Source	SGS335KBq	13 th August 2014
24	⁵⁷ Co	851 kBq	Point Source	1707-99-9	23 rd May 2014

- This source is used for all ²³⁸Pu contamination response testing unless stated otherwise on the calibration certificate.
- This source is used for all ²⁴¹Am contamination response testing unless stated otherwise on the calibration certificate.

Appendix B: Disclaimer

This document is issued, on the Client's behalf, by the Company under its General Conditions of Service available on request and accessible at

<http://www.sgs.com/en/Terms-and-Conditions/General-Conditions-of-Services-English.aspx>.

The Client's attention is drawn to the limitation of liability, indemnification and jurisdiction issues defined therein.

Any other holder of this document is advised that information contained hereon reflects the Company's findings at the time of its intervention only and within the limits of Client's instructions, if any. The Company's sole responsibility is to its Client and this document does not exonerate parties to a transaction from exercising all their rights and obligations under the transaction documents.

This report must not be reproduced, except in full.